### Food and Drug Administration, HHS

purpose, or the device is intended for lay use where the former intended use was by health care professionals only; or

(b) The modified device operates using a different fundamental scientific technology than that in use in the device before May 28, 1976; e.g., a surgical instrument cuts tissue with a laser beam rather than with a sharpened metal blade, or an in vitro diagnostic device detects or identifies inagents fectious by using deoxyribonucleic acid (DNA) probe or nucleic acid hybridization technology rather than culture or immunoassay technology.

[52 FR 32111, Aug. 25, 1987]

### Subpart B—Diagnostic Devices

#### §874.1050 Audiometer.

- (a) *Identification*. An audiometer or automated audiometer is an electroacoustic device that produces controlled levels of test tones and signals intended for use in conducting diagnostic hearing evaluations and assisting in the diagnosis of possible otologic disorders.
  - (b) Classification. Class II.

# §874.1060 Acoustic chamber for audiometric testing.

- (a) *Identification*. An acoustic chamber for audiometric testing is a room that is intended for use in conducting diagnostic hearing evaluations and that eliminates sound reflections and provides isolation from outside sounds.
- (b) *Classification.* Class I. The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter.
- [51 FR 40389, Nov. 6, 1986, as amended at 61 FR 1121, Jan. 16, 1996]

### §874.1070 Short increment sensitivity index (SISI) adapter.

(a) *Identification.* A short increment sensitivity index (SISI) adapter is a device used with an audiometer in diagnostic hearing evaluations. A SISI adapter provides short periodic sound pulses in specific small decibel increments that are intended to be superimposed on the audiometer's output tone frequency.

(b) *Classification.* Class I. [55 FR 48440, Nov. 20, 1990]

#### §874.1080 Audiometer calibration set.

- (a) Identification. An audiometer calibration set is an electronic reference device that is intended to calibrate an audiometer. It measures the sound frequency and intensity characteristics that emanate from an audiometer earphone. The device consists of an acoustic cavity of known volume, a sound level meter, a microphone with calibration traceable to the National Bureau of Standards, oscillators, frequency counters, microphone amplifiers, and a recorder. The device can measure selected audiometer test frequencies at a given intensity level, and selectable audiometer attenuation settings at a given test frequency.
- (b) *Classification*. Class I. The device is exempt from the premarket notification procedures in subpart E of part 807 of this chapter.
- $[51\ FR\ 40389,\ Nov.\ 6,\ 1986,\ as\ amended\ at\ 61\ FR\ 1121,\ Jan.\ 16,\ 1996]$

#### §874.1090 Auditory impedance tester.

- (a) Identification. An auditory impedance tester is a device that is intended to change the air pressure in the external auditory canal and measure and graph the mobility characteristics of the tympanic membrane to evaluate the functional condition of the middle ear. The device is used to determine abnormalities in the mobility of the tympanic membrane due to stiffness, flaccidity, or the presence of fluid in the middle ear cavity. The device is also used to measure the acoustic reflex threshold from contractions of the stapedial muscle, to monitor healing of tympanic membrane grafts stapedectomies, or to monitor followup treatment for inflammation of the middle ear.
  - (b) Classification. Class II.

# §874.1100 Earphone cushion for audiometric testing.

(a) *Identification*. An earphone cushion for audiometric testing is a device that is used to cover an audiometer earphone during audiometric testing to provide an acoustic coupling (sound